



EC-Conformity Declaration

For the following equipment :

Product Name: Switching Power Adaptor

Model Designation: GSM06X-Y (X=E, U) (Y=05, 06, 07, 09, 12, 15, 18, 24)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

Medical Directive (93/42/EEC) :

EN 60601-1:2006

TUV Certificate No : TA 50268288

Electromagnetic Compatibility Directive EN60601-1-2) :

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

	EN60601-1-2:2007, Group 1	CLASS B
Harmonic current	EN61000-3-2:2006+A1:2009+A2:2009,	CLASS A
Voltage flicker	EN61000-3-3:2008	
ESD air	EN61000-4-2:2008	Level 3 8KV
ESD contact	EN61000-4-2:2008	Level 3 6KV
RF field susceptibility	EN61000-4-3:2006+A1:2007+A2:2010	Level 2 3V/m
EFT bursts	EN61000-4-4:2004	Level 3 2KV/5KHz
Surge susceptibility	EN61000-4-5:2005	Level 3 1KV/Line-Line 2KV/Line-Earth
Conducted susceptibility	EN61000-4-6:2008	Level 2 3V
Magnetic field immunity	EN61000-4-8:2009	Level 2 3A/m
Voltage dip, interruption	EN61000-4-11:2004	>95% dip 0.5 periods, 60% dip 5 periods 30% dip 25 periods >95% interruptions 250 periods

Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

This Declaration is effective from serial number EJ xxxxxx B1401R

Person responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan (R.O.C.)

(Manufacturer Address)

Ted Cheng/Product Manager :

(Name / Position)

(Signature)

Taiwan

(Place)

Jan.15.2014

(Date)