

Eljintek, Inc.
Mr. Giedi Wang R&D Manager
-
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Date : 22.10.2019
Our ref. : ZTW1
Your ref.: GW-28082019-Co-licen

Ref : TA License

Type of Equipment : Switching Power Supply
Model Designation : See Certificate
Certificate No. : TA 50448649 0001
Report No. : 50219326 002

Dear Mr. Giedi Wang,

Enclosed please find above certification documents.
Please forward these originals to the certificate holder.

If you contact our office, please quote our reference above.

We thank you for your cooperation.

the certificate holder is: MEAN WELL Enterprises Co., Ltd.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428,
Taiwan, R. O. C.
Tel. (04) 2521-8888
Fax (04) 2521-8899

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

MEAN WELL Enterprises Co., Ltd.
Mr. Frank Lin Safety Engineer
-
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Date : 22.10.2019
Our ref. : ZTW1
Your ref.: GW-28082019-Co-licen

Ref : TA License

Type of Equipment : Switching Power Supply
Model Designation : See Certificate
Certificate No. : TA 50448649 0001
Report No. : 50219326 002

Dear Mr. Frank Lin,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

cc: Eljintek, Inc.

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428,
Taiwan, R. O. C.
Tel. (04) 2521-8888
Fax (04) 2521-8899

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50448649 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.: ZTW1- 50219326 002

Client Reference: GW-28082019-Co-licen

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: GSM06Eb (MW)	5
b stands for	: 05, 06, 07, 09, 12, 15, 18 or 24	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.18-0.09A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Ambient Temperature:	40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

6

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419073934

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

22/10/2019

Certificate



Certificate no.

TA 50448649 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1- 50219326 002

Client Reference: GW-28082019-Co-licen

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01,

Addition

Type Designation	: GEM06Ib (MW)	1
b stands for	: 05, 06, 07, 09, 12, 15, 18 or 24	1
Output Voltage	: see Appendix	
Output Current	: see Appendix	

Remark: This certificate is valid for maximum 5 years.

2

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419073934

TÜV Rheinland Taiwan Ltd

Signature



(Signature)
Dipl.-Ing. (FH) A. Klinker


Date of Issue

(day/mo/yr)

22/10/2019

Appendix to TÜV approved Certificate No.: TA 50448649

Certified Product : Switching Power Supply
 Report Number : 50219326 002

Type Designation: GSM06Eb, GEM06Ib (b = 05, 06, 07, 09, 12, 15, 18 or 24) ()

Output rating table:

Mean Well	Output Voltage (Vdc)	Output Current (A)
GSM06E05	5-6	1.2-1.0
GEM06I05		
GSM06E06 (SET: 6V, 1A) GSM06E07 (SET: 7.5V, 0.8A)	6-8	1.0-0.75
GEM06I06 (SET: 6V, 1A) GEM06I07 (SET: 7.5V, 0.8A)		
GSM06E09	8-11	0.75-0.55
GEM06I09		
GSM06E12	11-13	0.55-0.46
GEM06I12		
GSM06E15	13-16	0.46-0.38
GEM06I15		
GSM06E18	16-21	0.38-0.29
GEM06I18		
GSM06E24	21-24	0.29-0.26
GEM06I24		

Date: Oct. 22, 2019



Certification Body

Dipl.-Ing. (FH) A. Klinker