



### EC-Conformity Declaration

For the following equipment :

Desktop Power Supply  
(Product Name)

MES50x-y

x=A (for 3 pin inlet),C (for 3 pin small inlet) ; y=0~9

(Model Designation)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

#### Medical Device Directive (93/42/EEC) :

EN60601-1:1998    MES50A/C    TUV certificate No:TA50082469

#### Electromagnetic Compatibility (EN60601-1-2) :

Conducted emissions	EN55011	Class B	
Radiated emissions	EN55011	Class B	
Harmonic distortion	EN61000-3-2	Conforms	
Voltage flicker	EN61000-3-3	Conforms	
ESD air	EN61000-4-2	Level 4	8KV
ESD contact	EN61000-4-2	Level 3	6KV
RF field susceptibility	EN61000-4-3	Level 2	3V/m
EFT bursts	EN61000-4-4	Level 3	2KV/5KHz
Surge susceptibility	EN61000-4-5	Level 2	1KV/Line-Line
		Level 3	2KV/Line-Earth
Conducted susceptibility	EN61000-4-6	Level 2	3V/m
Magnetic field immunity	EN61000-4-8	Level 2	3A/m
Voltage dip, interruption	EN61000-4-11	Conforms	>95% dip 0.5 periods
			60% dip 5 periods
			30% dip 25 periods
			>95% interruptions 250 periods

Component power supply will be installed into a final equipment. Since EMC performance will be effected by complete installation, the final equipment manufacturers must re-conform EMC Directive on final equipment again.

#### Person responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.  
(Company Name)

No.28, Wu-Chuan 3rd Road, Wu Ku Ind. Park, Taipei Hsien, Taiwan  
(Company Address)

Gary Yen.  
(Name, Surname)

Products Manager  
(Position/Title)

Taiwan  
(Place)

Jun.11.2007  
(Date)

(Legal Signature) 2007.06.08