



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company and for below described product, based on

Technical Standard: MDD DIRECTIVE 93/42/EEC

General Information

Applicant: N/A

Product Description

EUT Description: Switching Power Supply

Brand Name: MEAN WELL

Model Number: MFM/MPM-15/20-X(X=3.3,5,12,15,24)

Measurement Standard

EN 55011:2009/A1:2010

EN 61000-3-2: 2014

EN 61000-3-3: 2013

EN 60601-1-2:2015

EN 61000-4-2:2009; EN 61000-4-3:2010

EN 61000-4-4:2012 ; EN 61000-4-5:2014 ; EN 61000-4-6:2014

EN 61000-4-8:2010 ; EN 61000-4-11:2004

Measurement Facilities

Kunshan BU. Name: **Compliance Certification Services Inc. Kunshan Laboratory**

No.10 Weiye Rd., Innovation park, Eco&Tec,

Development Zone, Kunshan City, Jiangsu, China

Tel: + 86-512-57355888/ FAX: +86-512-57370818

*This device has been tested and found to be in compliance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report with the number: **C170713E01-ET***

The test results indicated in this report are applicable only to the tested sample identified in this report.

Jeff Fang

Jeff.Fang /EMC Manager

Date: October 11, 2017