

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE**

Product  
Produit

Name and address of the applicant  
Nom et adresse du demandeur

Name and address of the manufacturer  
Nom et adresse du fabricant

Name and address of the factory  
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>eme</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Trademark (if any)  
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

Model / Type Ref.  
Ref. De type

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>eme</sup> page

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

As shown in the Test Report Ref. No. which forms part  
of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

**CERTIFICAT D'ESSAI OC**

Switching Power Supply

MEAN WELL ENTERPRISES CO LTD  
28 WUQUAN 3RD RD WUGU DIST NEW TAIPEI 248 TAIWAN

MEAN WELL ENTERPRISES CO LTD  
28 WUQUAN 3RD RD WUGU DIST NEW TAIPEI 248 TAIWAN

MEAN WELL ENTERPRISES CO., LTD.  
NO.28, WUQUAN 3RD RD., WUGU DIST., NEW TAIPEI CITY  
248,TAIWAN (R.O.C.)

Additional Information on page 2

See Page 2

MW, MEAN WELL



MSP-300-x  
See Page 2

National Differences specified in the CB Test Report.

Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

E227340-D9-CB-R1 issued on 2015-10-02

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/nbcnames](http://www.ul.com/nbcnames)

Date: 2015-10-14

Signature:

Jan-Erik Storgaard



Model Details:

MSP-300-x (x can be 3.3, 5, 7.5, 12, 15, 24, 36 or 48; x is to denote different output voltage)

Factories:

SUZHOU MEAN WELL TECHNOLOGY CO., LTD.

NO. 77, JIAN-MIN ROAD, DONG-QIAO, PAN-YANG IND. PARK, HUANG-DAI TOWN, XIANG-CHENG

DISTRICT, SUZHOU, JIANGSU 215152,

P. R. CHINA

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD.

2F, A BUILDING, YUEAN INDUSTRIAL PARK, HUANGCUN, DONGPU TOWN, TIANHE DISTRICT, GUANGZHOU,

GUANGDONG,

P.R. CHINA

Ratings:

Input: 100-240 V ac, 50/60 Hz, 4.5-1.6A

Output:

Model name	Output rating	Output power
MSP-300-3.3	3.3Vdc, 60A	198W
MSP-300-5	5Vdc, 60A	300W
MSP-300-7.5	7.5Vdc, 40A	300W
MSP-300-12	12Vdc, 27A	324W
MSP-300-15	15Vdc, 22A	330W
MSP-300-24	24Vdc, 14A	336W
MSP-300-36	36Vdc, 9A	324W
MSP-300-48	48Vdc, 7A	336W

Additional information:

Risk Management was not included in this investigation.

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2015-10-14

Signature:

Jan-Erik Storgaard