

# Verification of Compliance

Product Name : **Medical Type Switching Power Supply**  
Brand Name : **MEAN WELL**  
Model Number : **MSP-300-x (x=3.3,5,7.5,12,15,24,36,48)**  
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**  
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,  
Taiwan (R.O.C.)**  
Report Number : **S4C-M020-1110-171**  
Issue Date : **July 15, 2021**  
Applicable Standards :

**Emission :**

EN 55011: 2016/A2:2021 Group 1, Class B  
EN IEC 61000-3-2 : 2019  
EN 61000-3-3 : 2013/A1 :2019

**Immunity :**

EN 60601-1-2 : 2015 Home Healthcare Environment  
EN 61000-4-2 : 2009  
EN 61000-4-3 : 2006/A2 : 2010  
EN 61000-4-4 : 2012  
EN 61000-4-5 : 2014/A1:2017  
EN 61000-4-6 : 2014  
EN 61000-4-8 : 2010  
EN IEC 61000-4-11 : 2020



Based on the EMC Directive 2014/30/EU, MDR (Regulation (EU) 2017/745) and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



**Central Research Technology Co.**

11, Lane 41, Fushuen St., Jungshan Chiu,  
Taipei 104, Taiwan

Tel : 886-2-25984568

Fax : 886-2-25984546

(Jim Chien/Technical Manager)

Date: July 15, 2021

TAF 0905

FCC CAB Code TW1104, TW0019

NVLAP Lab Code 200575-0

ISED CAB Code TW0905

VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010

T-20009, G-10614