



Ref. Certif. No.

**DK-108090-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	Switching Power Supply
Name and address of the applicant	MEAN WELL ENTERPRISES CO LTD No 28 Wuquan 3rd Rd Wugu District New Taipei 24891 Taiwan
Name and address of the manufacturer	MEAN WELL ENTERPRISES CO LTD No 28 Wuquan 3rd Rd Wugu District New Taipei 24891 Taiwan
Name and address of the factory <small>Note: When more than one factory, please report on page 2</small>	MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd., Wugu District New Taipei City 24891, Taiwan
Ratings and principal characteristics	<input checked="" type="checkbox"/> Additional Information on page 2  See Page 2
Trademark / Brand (if any)	MW, MEAN WELL  
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	MSP-600-x See Page 2
Additional information (if necessary may also be reported on page 2)	<input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E227340-4789711340-1-Original issued on 2020-12-24

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-12-29

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-108090-UL**

**Model Details:**

MSP-600-x (x can be 3.3, 5, 7.5, 12, 15, 24, 36, or 48; x is to denote different output voltage)

**Factories:**

SuZhou MEAN WELL Technology Co., Ltd.

No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District Suzhou, Jiangsu 215152, China

MEAN WELL (GUANGZHOU) ELECTRONICS CO.,LTD HUADU BRANCH.

No.11 Jingu South Road, Huadong Town, Huadu District Guangzhou Guangdong 510890, China

YONGDEN TECHNOLOGY CORPORATION

345 MacArthur HighWay Tabang, Guiguinto Bulacan 3015 Philippines

**Ratings:**

Input: 100-240 V ac, 50/60 Hz, 8.5-5 A

**Output:**

Model name / Output rating / Output power

MSP-600-3.3 / 3.3Vdc, 120A / 396W

MSP-600-5 / 5Vdc, 120A / 600W

MSP-600-7.5 / 7.5Vdc, 80A / 600W

MSP-600-12 / 12Vdc, 53A / 636W

MSP-600-15 / 15Vdc, 43A / 645W

MSP-600-24 / 24Vdc, 27A / 648W

MSP-600-36 / 36Vdc, 17.5A / 630W

MSP-600-48 / 48Vdc, 13A / 624W

**Additional Information:**

Additionally evaluated to EN 60601-1:2006/A1:2013/A12:2014.

National Differences specified in the CB Test Report.

Risk Management was not included in this investigation.

**Additional information (if necessary)**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-12-29

Signature:

Jan-Erik Storgaard