

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Mode Power Supply
Name and address of the applicant	MEAN WELL Enterprises Co.,Ltd. No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co.,Ltd. No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan
Name and address of the factory	(See appendix for factories information)
Ratings and principal characteristics	AC 100-240V; 50/60Hz; 1.8-1.1A; Class I; for output ratings refer to test report no. 11026836 001 page 5
Trademark (if any)	Trademark of MEAN WELL
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	RPx-75y (x = S, D, T to denote the different sets of DC output; y = -3.3, -5, -12, -15, -24, -36, -48, A, B, C, D, 03 to denote different output voltage)
Additional information (if necessary may also be reported on page 2)	-change 'Test item description'; for details see test report; -see also test report ref. no. 11026836 001; -supersede previously issued certificate DE 2-015488
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	11026836 002

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 - 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 05.12.2017

Signature:

Dipl.-Ing. (FH) A. Klinker

1. MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan
2. Suzhou MEAN WELL Technology
Co., Ltd. No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng Dist.
Suzhou, 215152 Jiangsu, China
3. MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road
Huadong Town, Huadu District
Guangzhou, 510890 Guangdong, China

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date: 05.12.2017

Signature:

Dipl.-Ing. (FH) A. Klinker

