



QQHM2.E227340
Power Supplies, Medical and Dental - Component

Page Bottom

Power Supplies, Medical and Dental - Component

See General Information for Power Supplies, Medical and Dental - Component

MEAN WELL ENTERPRISES CO LTD

E227340

28 WU-CHUAN 3RD RD
 WUGU INDUSTRIAL PARK
 TAIPEI HSIEN, 248 TAIWAN

| Model No. | Rated Input | | SC | Output | | | OC | SP | EP | FC | GC |
|---------------|-------------|-------|----|--------|-------|------------|----|------|-----|----|----|
| | V | Hz | | Max V | Max A | Max VA (b) | | | | | |
| MPS-30-48 (a) | 100-240 | 50/60 | 0 | 48 | 0.6 | 28.8 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-27 (a) | 100-240 | 50/60 | 0 | 27 | 1.1 | 29.7 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-24 (a) | 100-240 | 50/60 | 0 | 24 | 1.2 | 28.8 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-15 (a) | 100-240 | 50/60 | 0 | 15 | 2 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-12 (a) | 100-240 | 50/60 | 0 | 12 | 2.5 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-5 (a) | 100-240 | 50/60 | 0 | 5 | 6 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-45-3.3 | 100-240 | 50/60 | 0 | 3.3R | 8R | 26.4R | 9 | 2601 | — | 0 | 0 |
| MPS-45-5 | 100-240 | 50/60 | 0 | 5R | 8R | 40.0R | 9 | 2601 | — | 0 | 0 |
| MPS-45-7.5 | 100-240 | 50/60 | 0 | 7.5R | 5.5R | 41.3R | 9 | 2601 | — | 0 | 0 |
| MPS-45-12 | 100-240 | 50/60 | 0 | 12R | 3.7R | 44.4R | 9 | 2601 | — | 0 | 0 |
| MPS-45-13.5 | 100-240 | 50/60 | 0 | 13.5R | 3.3R | 44.6R | 9 | 2601 | — | 0 | 0 |
| MPS-45-15 | 100-240 | 50/60 | 0 | 15R | 3R | 45.0R | 9 | 2601 | — | 0 | 0 |
| MPS-45-24 | 100-240 | 50/60 | 0 | 24R | 1.9R | 45.6R | 9 | 2601 | — | 0 | 0 |
| MPS-45-27 | 100-240 | 50/60 | 0 | 27R | 1.7R | 45.9R | 9 | 2601 | — | 0 | 0 |
| MPS-45-48 | 100-240 | 50/60 | 0 | 48R | 1R | 48.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-3.3 | 100-240 | 50/60 | 0 | 3.3R | 12R | 39.6R | 9 | 2601 | — | 0 | 0 |
| MPS-65-5 | 100-240 | 50/60 | 0 | 5R | 12R | 60.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-7.5 | 100-240 | 50/60 | 0 | 7.5R | 8R | 60.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-12 | 100-240 | 50/60 | 0 | 12R | 5.2R | 62.4R | 9 | 2601 | — | 0 | 0 |

| | | | | | | | | | | | |
|--|---------|-------|---|-------|-----------|-------|---|---------|----|---|---|
| MPS-65-13.5 | 100-240 | 50/60 | 0 | 13.5R | 4.7R | 63.5R | 9 | 2601 | — | 0 | 0 |
| MPS-65-15 | 100-240 | 50/60 | 0 | 15R | 4.2R | 63.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-24 | 100-240 | 50/60 | 0 | 24R | 2.7R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPS-65-27 | 100-240 | 50/60 | 0 | 27R | 2.4R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPS-65-48 | 100-240 | 50/60 | 0 | 48R | 1.35R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPD-45A | 100-240 | 50/60 | 0 | 5R | 3.2R | 16.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| MPD-45B | 100-240 | 50/60 | 0 | 5R | 3.2R | 16.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 24R | 1.2R | 28.8R | | | | | |
| MPD-65A | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.8R | 33.6R | | | | | |
| MPD-65B | 100-240 | 50/60 | 0 | 5R | 3.5R | 17.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 24R | 2R | 48R | | | | | |
| MPT-45A | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| | | | | 12R | 2R | 24R | | | | | |
| MPT-45B | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| | | | | -12R | 0.3R | 3.6R | | | | | |
| MPT-45C | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 15R | 1.6R | 24R | | | | | |
| | | | | -15R | 0.3R | 4.5R | | | | | |
| MPT-65A | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.5R | 30R | | | | | |
| | | | | -5R | 0.5R | 2.5R | | | | | |
| MPT-65B | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.5R | 30R | | | | | |
| | | | | -12R | 0.5R | 6R | | | | | |
| MPT-65C | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 15R | 2R | 30R | | | | | |
| | | | | -15R | 0.5R | 7.5R | | | | | |
| MES30A-0, MES30B-0, MES30C-0, MES30D-0-U | 100-240 | 50-60 | 0 | 3-5 | 5.0 | 25 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-1, MES30B-1, MES30C-1, MES30D-1-U | 100-240 | 50-60 | 0 | 5-6 | 5.0-4.17 | 25 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-1-1, MES30B-1-1, MES30C-1-1, MES30D-1-1-U | 100-240 | 50-60 | 0 | 6-8 | 4.17-3.13 | 25 | 9 | 60601-1 | 2T | 0 | 1 |

| | | | | | | | | | | | |
|--|---------|-------|---|----------|-----------|----|---|---------|-----|---|---|
| MES30A-2, MES30B-2, MES30C-2, MES30D-2-U | 100-240 | 50-60 | 0 | 8-11 | 3.75-2.73 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-3, MES30B-3, MES30C-3, MES30D-3-U | 100-240 | 50-60 | 0 | 11-13 | 2.73-2.30 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-4, MES30B-4, MES30C-4, MES30D-4-U | 100-240 | 50-60 | 0 | 13-16 | 2.30-1.88 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-5, MES30B-5, MES30C-5, MES30D-5-U | 100-240 | 50-60 | 0 | 16-21 | 1.88-1.43 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-6, MES30B-6, MES30C-6, MES30D-6-U | 100-240 | 50-60 | 0 | 21-27 | 1.43-1.11 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-7, MES30B-7, MES30C-7, MES30D-7-U | 100-240 | 50-60 | 0 | 27-33 | 1.11-0.90 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-8, MES30B-8, MES30C-8, MES30D-8-U | 100-240 | 50-60 | 0 | 33-48 | 0.9-0.63 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MPS-120-3.3 | 100-240 | 50/60 | 0 | +3.3 (a) | 24 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-5 | 100-240 | 50/60 | 0 | +5 (a) | 22 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-12 | 100-240 | 50/60 | 0 | +12 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-15 | 100-240 | 50/60 | 0 | +15 (a) | 8 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-24 | 100-240 | 50/60 | 0 | +24 (a) | 5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-48 | 100-240 | 50/60 | 0 | +48 (a) | 2.5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPD-120A | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPD-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.9 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120A | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.8 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | - 5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.4 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120C | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 4 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120D | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |

| | | | | | | | | | | | |
|-------------|---------|-------|---|----------|---------|---|---|---------|-----|---|---|
| MPQ-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120C | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 3.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120D | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 1 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.1 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120E | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 3 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-200-3.3 | 100-240 | 50/60 | 0 | +3.3 (a) | 40 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-5 | 100-240 | 50/60 | 0 | +5 (a) | 40 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-7.5 | 100-240 | 50/60 | 0 | +7.5 (a) | 26.6 | — | — | 60601-1 | 20B | 0 | 1 |
| MPS-200-12 | 100-240 | 50/60 | 0 | +12 (a) | 16.7 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-15 | 100-240 | 50/60 | 0 | +15 (a) | 13.4 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-24 | 100-240 | 50/60 | 0 | +24 (a) | 8.4 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-48 | 100-240 | 50/60 | 0 | +48 (a) | 4.2 | | | 60601-1 | 20B | 0 | 1 |
| MPD-200A | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 8 | | | | | | |
| MPD-200B | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 4 | | | | | | |
| MPT-200A | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 7.5 | | | | | | |

| | | | | | | | | | | | | |
|-----------------------|---------|-------|---|--------|---------|----------|---|---------|-----|---|---|--|
| | | | | | -5 (a) | 2 | | | | | | |
| MPT-200B | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +12 (a) | 6 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPT-200C | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +15 (a) | 4.7 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| MPT-200D | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +24 (a) | 3 | | | | | | |
| | | | | | +12 (a) | 2 | | | | | | |
| MPQ-200B | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +12 (a) | 7 | | | | | | |
| | | | | | -5 (a) | 2 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPQ-200C | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +15 (a) | 5 | | | | | | |
| | | | | | -5 (a) | 2 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| MPQ-200D | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +12 (a) | 2 | | | | | | |
| | | | | | +24 (a) | 3 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPQ-200F | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +15 (a) | 2 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| | | | | | +24 (a) | 2.7 | | | | | | |
| PM-10-3.3, NFM-10-3.3 | 100-240 | 50/60 | 0 | 3.3(a) | 2.5(a) | 8.25(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-5, NFM-10-5 | 100-240 | 50/60 | 0 | 5(a) | 2.0(a) | 10(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-12, NFM-10-12 | 100-240 | 50/60 | 0 | 12(a) | 0.85(a) | 10.2(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-15, NFM-10-15 | 100-240 | 50/60 | 0 | 15(a) | 0.67(a) | 10.05(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-24, NFM-10-24 | 100-240 | 50/60 | 0 | 24(a) | 0.42(a) | 10.08(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-15-3.3, NFM-15-3.3 | 100-240 | 50/60 | 0 | 3.3(a) | 3.5(a) | 11.55(a) | 9 | 60601-1 | 20B | 0 | 2 | |

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | | | | | | | | | |
|-----------------------------|---------|-------|---|---------|-----------|----------|---|---------|-----|---|---|
| PM-15-5, NFM-15-5 | 100-240 | 50/60 | 0 | 5(a) | 3.0(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-12, NFM-15-12 | 100-240 | 50/60 | 0 | 12(a) | 1.25(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-15, NFM-15-15 | 100-240 | 50/60 | 0 | 15(a) | 1.0(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-24, NFM-15-24 | 100-240 | 50/60 | 0 | 24(a) | 0.63(a) | 15.12(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-3.3, NFM-05-3.3 | 100-240 | 50/60 | 0 | 24(a) | 1.25(a) | 4.125(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-5, NFM-05-5 | 100-240 | 50/60 | 0 | 24(a) | 1.0(a) | 5.0(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-12, NFM-05-12 | 100-240 | 50/60 | 0 | 24(a) | 0.42(a) | 5.04(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-15, NFM-05-15 | 100-240 | 50/60 | 0 | 24(a) | 0.33(a) | 4.95(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-24, NFM-05-24 | 100-240 | 50/60 | 0 | 24(a) | 0.23(a) | 5.52(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-20-3.3(a), NFM-20-3.3(a) | 100-240 | 50/60 | 0 | 3.3 | 4.5 | 14.85 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-5(a), NFM-20-5(a) | 100-240 | 50/60 | 0 | 5.0 | 4.4 | 22.0 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-12(a), NFM-20-12(a) | 100-240 | 50/60 | 0 | 12.0 | 1.8 | 21.6 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-15(a), NFM-20-15(a) | 100-240 | 50/60 | 0 | 15.0 | 1.4 | 21.0 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-24(a), NFM-20-24(a) | 100-240 | 50/60 | 0 | 24.0 | 0.92 | 22.08 | 9 | 60601-1 | 20B | 0 | 1 |
| MES50X-0(a) (c) | 100-240 | 50/60 | 0 | 3-5 | 7.5 | 37.5 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-1(a) (c) | 100-240 | 50/60 | 0 | 5-6 | 7.50-6.25 | 37.5 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-1-1(a) (c) | 100-240 | 50/60 | 0 | 6-8 | 6.66-5.00 | 40 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-2(a) (c) | 100-240 | 50/60 | 0 | 8-11 | 5.62-4.09 | 45 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-3(a) (c) | 100-240 | 50/60 | 0 | 11-13 | 4.54-3.84 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-4(a) (c) | 100-240 | 50/60 | 0 | 13-16 | 3.84-3.12 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-5(a) (c) | 100-240 | 50/60 | 0 | 16-21 | 3.12-2.38 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-6(a) (c) | 100-240 | 50/60 | 0 | 21-27 | 2.38-1.85 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-7(a) (c) | 100-240 | 50/60 | 0 | 27-33 | 1.85-1.51 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-8(a) (c) | 100-240 | 50/60 | 0 | 33-48 | 1.51-1.04 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| RPS-60-3.3 | 100-240 | 50/60 | 0 | +3.3(a) | 10(a) | 33(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-5 | 100-240 | 50/60 | 0 | +5(a) | 10(a) | 50(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-12 | 100-240 | 50/60 | 0 | +12(a) | 5(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-15 | 100-240 | 50/60 | 0 | +15(a) | 4(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-24 | 100-240 | 50/60 | 0 | +24(a) | 2.5(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |

| | | | | | | | | | | | |
|----------|---------|-------|---|---------|--------|---------|---|---------|-----|---|---|
| RPT-75B | 100-240 | 50/60 | 0 | +5(a) | 6(a) | 72.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12(a) | 3(a) | | | | | | |
| | | | | -12(a) | 0.5(a) | | | | | | |
| RPT-75C | 100-240 | 50/60 | 0 | +5(a) | 6(a) | 72.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15(a) | 2.3(a) | | | | | | |
| | | | | -15(a) | 0.5(a) | | | | | | |
| RPT-75D | 100-240 | 50/60 | 0 | +5(a) | 5(a) | 73.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24(a) | 1.5(a) | | | | | | |
| | | | | +12(a) | 1(a) | | | | | | |
| RPT-7503 | 100-240 | 50/60 | 0 | +3.3(a) | 6(a) | 61.8(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +5(a) | 6(a) | | | | | | |
| | | | | +12(a) | 1(a) | | | | | | |

(a) Output values are rated, not maximum.

(b) Ratings in W not VA.

(c) Where X may be A or C.

Marking: Company name and model designation.

Last Updated on 2006-12-27

[Questions?](#)

[Notice of Disclaimer](#)

[Page Top](#)

Copyright © 2007 Underwriters Laboratories Inc.®

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Listed and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Designs and/or Listings (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from the Online Certifications Directory with permission from Underwriters Laboratories Inc." must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following format: "Copyright © 2007 Underwriters Laboratories Inc.®"

An independent organization working for a safer world with integrity, precision and knowledge.





QQHM8.E227340

Power Supplies, Medical and Dental Certified for Canada - Component

Page Bottom

Power Supplies, Medical and Dental Certified for Canada - Component

See General Information for Power Supplies, Medical and Dental Certified for Canada - Component

MEAN WELL ENTERPRISES CO LTD

E227340

28 WU-CHUAN 3RD RD
WUGU INDUSTRIAL PARK
TAIPEI HSIEN, 248 TAIWAN

| Model No. | Rated Input | | SC | Output | | | OC | SP | EP | FC | GC |
|---------------|-------------|-------|----|--------|-------|------------|----|------|-----|----|----|
| | V | Hz | | Max V | Max A | Max VA (b) | | | | | |
| MPS-30-48 (a) | 100-240 | 50/60 | 0 | 48 | 0.6 | 28.8 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-27 (a) | 100-240 | 50/60 | 0 | 27 | 1.1 | 29.7 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-24 (a) | 100-240 | 50/60 | 0 | 24 | 1.2 | 28.8 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-15 (a) | 100-240 | 50/60 | 0 | 15 | 2 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-12 (a) | 100-240 | 50/60 | 0 | 12 | 2.5 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-30-5 (a) | 100-240 | 50/60 | 0 | 5 | 6 | 30 | 9 | 2601 | 20B | 0 | 0 |
| MPS-45-3.3 | 100-240 | 50/60 | 0 | 3.3R | 8R | 26.4R | 9 | 2601 | — | 0 | 0 |
| MPS-45-5 | 100-240 | 50/60 | 0 | 5R | 8R | 40.0R | 9 | 2601 | — | 0 | 0 |
| MPS-45-7.5 | 100-240 | 50/60 | 0 | 7.5R | 5.5R | 41.3R | 9 | 2601 | — | 0 | 0 |
| MPS-45-12 | 100-240 | 50/60 | 0 | 12R | 3.7R | 44.4R | 9 | 2601 | — | 0 | 0 |
| MPS-45-13.5 | 100-240 | 50/60 | 0 | 13.5R | 3.3R | 44.6R | 9 | 2601 | — | 0 | 0 |
| MPS-45-15 | 100-240 | 50/60 | 0 | 15R | 3R | 45.0R | 9 | 2601 | — | 0 | 0 |
| MPS-45-24 | 100-240 | 50/60 | 0 | 24R | 1.9R | 45.6R | 9 | 2601 | — | 0 | 0 |
| MPS-45-27 | 100-240 | 50/60 | 0 | 27R | 1.7R | 45.9R | 9 | 2601 | — | 0 | 0 |
| MPS-45-48 | 100-240 | 50/60 | 0 | 48R | 1R | 48.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-3.3 | 100-240 | 50/60 | 0 | 3.3R | 12R | 39.6R | 9 | 2601 | — | 0 | 0 |
| MPS-65-5 | 100-240 | 50/60 | 0 | 5R | 12R | 60.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-7.5 | 100-240 | 50/60 | 0 | 7.5R | 8R | 60.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-12 | 100-240 | 50/60 | 0 | 12R | 5.2R | 62.4R | 9 | 2601 | — | 0 | 0 |

| | | | | | | | | | | | |
|--|---------|-------|---|-------|-----------|-------|---|---------|----|---|---|
| MPS-65-13.5 | 100-240 | 50/60 | 0 | 13.5R | 4.7R | 63.5R | 9 | 2601 | — | 0 | 0 |
| MPS-65-15 | 100-240 | 50/60 | 0 | 15R | 4.2R | 63.0R | 9 | 2601 | — | 0 | 0 |
| MPS-65-24 | 100-240 | 50/60 | 0 | 24R | 2.7R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPS-65-27 | 100-240 | 50/60 | 0 | 27R | 2.4R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPS-65-48 | 100-240 | 50/60 | 0 | 48R | 1.35R | 64.8R | 9 | 2601 | — | 0 | 0 |
| MPD-45A | 100-240 | 50/60 | 0 | 5R | 3.2R | 16.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| MPD-45B | 100-240 | 50/60 | 0 | 5R | 3.2R | 16.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 24R | 1.2R | 28.8R | | | | | |
| MPD-65A | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.8R | 33.6R | | | | | |
| MPD-65B | 100-240 | 50/60 | 0 | 5R | 3.5R | 17.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 24R | 2R | 48R | | | | | |
| MPT-45A | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| | | | | 12R | 2R | 24R | | | | | |
| MPT-45B | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2R | 24R | | | | | |
| | | | | -12R | 0.3R | 3.6R | | | | | |
| MPT-45C | 100-240 | 50/60 | 0 | 5R | 3R | 15.0R | 9 | 2601 | — | 0 | 0 |
| | | | | 15R | 1.6R | 24R | | | | | |
| | | | | -15R | 0.3R | 4.5R | | | | | |
| MPT-65A | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.5R | 30R | | | | | |
| | | | | -5R | 0.5R | 2.5R | | | | | |
| MPT-65B | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 12R | 2.5R | 30R | | | | | |
| | | | | -12R | 0.5R | 6R | | | | | |
| MPT-65C | 100-240 | 50/60 | 0 | 5R | 5.5R | 27.5R | 9 | 2601 | — | 0 | 0 |
| | | | | 15R | 2R | 30R | | | | | |
| | | | | -15R | 0.5R | 7.5R | | | | | |
| MES30A-0, MES30B-0, MES30C-0, MES30D-0-U | 100-240 | 50-60 | 0 | 3-5 | 5.0 | 25 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-1, MES30B-1, MES30C-1, MES30D-1-U | 100-240 | 50-60 | 0 | 5-6 | 5.0-4.17 | 25 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-1-1, MES30B-1-1, MES30C-1-1, MES30D-1-1-U | 100-240 | 50-60 | 0 | 6-8 | 4.17-3.13 | 25 | 9 | 60601-1 | 2T | 0 | 1 |

| | | | | | | | | | | | |
|--|---------|-------|---|----------|-----------|----|---|---------|-----|---|---|
| MES30A-2, MES30B-2, MES30C-2, MES30D-2-U | 100-240 | 50-60 | 0 | 8-11 | 3.75-2.73 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-3, MES30B-3, MES30C-3, MES30D-3-U | 100-240 | 50-60 | 0 | 11-13 | 2.73-2.30 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-4, MES30B-4, MES30C-4, MES30D-4-U | 100-240 | 50-60 | 0 | 13-16 | 2.30-1.88 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-5, MES30B-5, MES30C-5, MES30D-5-U | 100-240 | 50-60 | 0 | 16-21 | 1.88-1.43 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-6, MES30B-6, MES30C-6, MES30D-6-U | 100-240 | 50-60 | 0 | 21-27 | 1.43-1.11 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-7, MES30B-7, MES30C-7, MES30D-7-U | 100-240 | 50-60 | 0 | 27-33 | 1.11-0.90 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MES30A-8, MES30B-8, MES30C-8, MES30D-8-U | 100-240 | 50-60 | 0 | 33-48 | 0.9-0.63 | 30 | 9 | 60601-1 | 2T | 0 | 1 |
| MPS-120-3.3 | 100-240 | 50/60 | 0 | +3.3 (a) | 24 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-5 | 100-240 | 50/60 | 0 | +5 (a) | 22 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-12 | 100-240 | 50/60 | 0 | +12 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-15 | 100-240 | 50/60 | 0 | +15 (a) | 8 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-24 | 100-240 | 50/60 | 0 | +24 (a) | 5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-120-48 | 100-240 | 50/60 | 0 | +48 (a) | 2.5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPD-120A | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 5 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPD-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.9 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120A | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.8 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | - 5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.4 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120C | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 4 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPT-120D | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |

| | | | | | | | | | | | |
|-------------|---------|-------|---|----------|---------|---|---|---------|-----|---|---|
| MPQ-120B | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 4.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120C | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 3.2 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -5 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120D | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 1 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 2.1 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | -12 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPQ-120E | 100-240 | 50/60 | 0 | +5 (a) | 10 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 3 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 0.6 (a) | - | 9 | 60601-1 | 20B | 0 | 1 |
| MPS-200-3.3 | 100-240 | 50/60 | 0 | +3.3 (a) | 40 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-5 | 100-240 | 50/60 | 0 | +5 (a) | 40 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-7.5 | 100-240 | 50/60 | 0 | +7.5 (a) | 26.6 | — | — | 60601-1 | 20B | 0 | 1 |
| MPS-200-12 | 100-240 | 50/60 | 0 | +12 (a) | 16.7 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-15 | 100-240 | 50/60 | 0 | +15 (a) | 13.4 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-24 | 100-240 | 50/60 | 0 | +24 (a) | 8.4 | | | 60601-1 | 20B | 0 | 1 |
| MPS-200-48 | 100-240 | 50/60 | 0 | +48 (a) | 4.2 | | | 60601-1 | 20B | 0 | 1 |
| MPD-200A | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 8 | | | | | | |
| MPD-200B | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +24 (a) | 4 | | | | | | |
| MPT-200A | 100-240 | 50/60 | 0 | +5 (a) | 20 | | | 60601-1 | 20B | 0 | 1 |
| | | | | +12 (a) | 7.5 | | | | | | |

| | | | | | | | | | | | | |
|-----------------------|---------|-------|---|--------|---------|----------|---|---------|-----|---|---|--|
| | | | | | -5 (a) | 2 | | | | | | |
| MPT-200B | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +12 (a) | 6 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPT-200C | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +15 (a) | 4.7 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| MPT-200D | 100-240 | 50/60 | 0 | | +5 (a) | 20 | | | | | | |
| | | | | | +24 (a) | 3 | | | | | | |
| | | | | | +12 (a) | 2 | | | | | | |
| MPQ-200B | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +12 (a) | 7 | | | | | | |
| | | | | | -5 (a) | 2 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPQ-200C | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +15 (a) | 5 | | | | | | |
| | | | | | -5 (a) | 2 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| MPQ-200D | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +12 (a) | 2 | | | | | | |
| | | | | | +24 (a) | 3 | | | | | | |
| | | | | | -12 (a) | 2 | | | | | | |
| MPQ-200F | 100-240 | 50/60 | 0 | | +5 (a) | 15 | | | | | | |
| | | | | | +15 (a) | 2 | | | | | | |
| | | | | | -15 (a) | 2 | | | | | | |
| | | | | | +24 (a) | 2.7 | | | | | | |
| PM-10-3.3, NFM-10-3.3 | 100-240 | 50/60 | 0 | 3.3(a) | 2.5(a) | 8.25(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-5, NFM-10-5 | 100-240 | 50/60 | 0 | 5(a) | 2.0(a) | 10(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-12, NFM-10-12 | 100-240 | 50/60 | 0 | 12(a) | 0.85(a) | 10.2(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-15, NFM-10-15 | 100-240 | 50/60 | 0 | 15(a) | 0.67(a) | 10.05(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-10-24, NFM-10-24 | 100-240 | 50/60 | 0 | 24(a) | 0.42(a) | 10.08(a) | 9 | 60601-1 | 20B | 0 | 2 | |
| PM-15-3.3, NFM-15-3.3 | 100-240 | 50/60 | 0 | 3.3(a) | 3.5(a) | 11.55(a) | 9 | 60601-1 | 20B | 0 | 2 | |

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | |
|---------|-----|---|---|
| 60601-1 | 20B | 0 | 1 |
|---------|-----|---|---|

| | | | | | | | | | | | |
|-----------------------------|---------|-------|---|---------|-----------|----------|---|---------|-----|---|---|
| PM-15-5, NFM-15-5 | 100-240 | 50/60 | 0 | 5(a) | 3.0(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-12, NFM-15-12 | 100-240 | 50/60 | 0 | 12(a) | 1.25(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-15, NFM-15-15 | 100-240 | 50/60 | 0 | 15(a) | 1.0(a) | 15(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-15-24, NFM-15-24 | 100-240 | 50/60 | 0 | 24(a) | 0.63(a) | 15.12(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-3.3, NFM-05-3.3 | 100-240 | 50/60 | 0 | 24(a) | 1.25(a) | 4.125(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-5, NFM-05-5 | 100-240 | 50/60 | 0 | 24(a) | 1.0(a) | 5.0(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-12, NFM-05-12 | 100-240 | 50/60 | 0 | 24(a) | 0.42(a) | 5.04(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-15, NFM-05-15 | 100-240 | 50/60 | 0 | 24(a) | 0.33(a) | 4.95(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-05-24, NFM-05-24 | 100-240 | 50/60 | 0 | 24(a) | 0.23(a) | 5.52(a) | 9 | 60601-1 | 20B | 0 | 2 |
| PM-20-3.3(a), NFM-20-3.3(a) | 100-240 | 50/60 | 0 | 3.3 | 4.5 | 14.85 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-5(a), NFM-20-5(a) | 100-240 | 50/60 | 0 | 5.0 | 4.4 | 22.0 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-12(a), NFM-20-12(a) | 100-240 | 50/60 | 0 | 12.0 | 1.8 | 21.6 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-15(a), NFM-20-15(a) | 100-240 | 50/60 | 0 | 15.0 | 1.4 | 21.0 | 9 | 60601-1 | 20B | 0 | 1 |
| PM-20-24(a), NFM-20-24(a) | 100-240 | 50/60 | 0 | 24.0 | 0.92 | 22.08 | 9 | 60601-1 | 20B | 0 | 1 |
| MES50X-0(a) (c) | 100-240 | 50/60 | 0 | 3-5 | 7.5 | 37.5 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-1(a) (c) | 100-240 | 50/60 | 0 | 5-6 | 7.50-6.25 | 37.5 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-1-1(a) (c) | 100-240 | 50/60 | 0 | 6-8 | 6.66-5.00 | 40 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-2(a) (c) | 100-240 | 50/60 | 0 | 8-11 | 5.62-4.09 | 45 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-3(a) (c) | 100-240 | 50/60 | 0 | 11-13 | 4.54-3.84 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-4(a) (c) | 100-240 | 50/60 | 0 | 13-16 | 3.84-3.12 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-5(a) (c) | 100-240 | 50/60 | 0 | 16-21 | 3.12-2.38 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-6(a) (c) | 100-240 | 50/60 | 0 | 21-27 | 2.38-1.85 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-7(a) (c) | 100-240 | 50/60 | 0 | 27-33 | 1.85-1.51 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| MES50X-8(a) (c) | 100-240 | 50/60 | 0 | 33-48 | 1.51-1.04 | 50 | 9 | 60601-1 | 20B | 8 | 1 |
| RPS-60-3.3 | 100-240 | 50/60 | 0 | +3.3(a) | 10(a) | 33(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-5 | 100-240 | 50/60 | 0 | +5(a) | 10(a) | 50(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-12 | 100-240 | 50/60 | 0 | +12(a) | 5(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-15 | 100-240 | 50/60 | 0 | +15(a) | 4(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |
| RPS-60-24 | 100-240 | 50/60 | 0 | +24(a) | 2.5(a) | 60(a) | 9 | 60601-1 | 20B | 0 | 2 |

| | | | | | | | | | | | |
|----------|---------|-------|---|---------|--------|---------|---|---------|-----|---|---|
| RPT-75B | 100-240 | 50/60 | 0 | +5(a) | 6(a) | 72.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +12(a) | 3(a) | | | | | | |
| | | | | -12(a) | 0.5(a) | | | | | | |
| RPT-75C | 100-240 | 50/60 | 0 | +5(a) | 6(a) | 72.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +15(a) | 2.3(a) | | | | | | |
| | | | | -15(a) | 0.5(a) | | | | | | |
| RPT-75D | 100-240 | 50/60 | 0 | +5(a) | 5(a) | 73.0(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +24(a) | 1.5(a) | | | | | | |
| | | | | +12(a) | 1(a) | | | | | | |
| RPT-7503 | 100-240 | 50/60 | 0 | +3.3(a) | 6(a) | 61.8(a) | 9 | 60601-1 | 20B | 0 | 1 |
| | | | | +5(a) | 6(a) | | | | | | |
| | | | | +12(a) | 1(a) | | | | | | |

(a) Output values are rated, not maximum.

(b) Ratings in W not VA.

(c) Where X may be A or C.



Marking: Company name, catalog designation and Recognized Component Mark for Canada

Last Updated on 2006-12-27

[Questions?](#)

[Notice of Disclaimer](#)

[Page Top](#)

Copyright © 2007 Underwriters Laboratories Inc.®

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Listed and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Designs and/or Listings (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from the Online Certifications Directory with permission from Underwriters Laboratories Inc." must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following format: "Copyright © 2007 Underwriters Laboratories Inc.®"

An independent organization working for a safer world with integrity, precision and knowledge.

